

ePM 10/12/15

Patient Monitor

Data Sheet



Physical Specifications

| | |
|--|---|
| Weight | ePM 10: 3.2 Kg ePM 12: 3.4 Kg ePM 15: 4.9 Kg (Standard configuration, excluding recorder, battery and accessories.) |
| Size | ePM 10: 271 x 226 x 173 mm ePM 12: 312 x 258 x 174 mm ePM 15: 397 x 293 x 181 mm |
| Display screen | Capacitive screen, support multi-touch operation. ePM 10: 10.1-inch, 1280 x 800 pixels ePM 12: 12.1-inch, 1280 x 800 pixels ePM 15: 15.6-inch, 1366 x 768 pixels |
| Display channel | ePM 10: Up to 8 waveform channels ePM 12: Up to 10 waveform channels ePM 15: Up to 12 waveform channels |
| ePM 10 main unit complies with the requirements of 6.3.4.3, EN1789 | |
| Drop test: | 0.75m for each of the 6 surfaces (ePM 10) |

ECG

Meet standards of IEC 60601-2-27 and IEC 60601-2-25.

| | |
|---|--|
| Lead set | 3-lead: I, II, III 5-lead: I, II, III, aVR, aVL, aVF, V ** 6-lead: I, II, III, aVR, aVL, aVF, Va, Vb 12-lead: I, II, III, aVR, aVL, aVF, V1 to V6 |
| Automatic 3/5/6/12 - lead recognition. | |
| Input signal range | ± 10 mV (p-p) |
| Electrode offset potential tolerance | ± 800 mV |
| Sweep speed | 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s |
| Gain | x 0.125, x 0.25, x 0.5, x 1, x 2, x 4, auto |
| Waveform format | Standard, Cabrera |
| Bandwidth | Diagnostic mode: 0.05 to 150 Hz Monitor mode: 0.5 to 40 Hz Surgical mode: 1 to 20 Hz ST mode: 0.05 to 40 Hz |
| CMRR | Diagnostic mode: > 90 dB Monitor, Surgical, ST mode: > 105 dB |
| Pace detection | Amplitude: ± 2 mV to ± 700 mV Width: 0.1 to 2 ms Rise time: 10 to 100 µs |
| Defib. protection | Withstand 5000V (360J) defibrillation |
| Recovery time | <5 s |
| Provides glasgow resting 12-lead ECG algorithm, and 12-lead ECG is not available for ePM 10 | |

Heart Rate

| | |
|---------------|--|
| HR rang | Adult: 15 to 300 bpm Pediatric/Neonate: 15 to 350 bpm |
| HR accuracy | ± 1 bpm or ± 1%, whichever is greater. |
| HR resolution | 1 bpm |

Arrhythmia Analysis

Intended use for adult, pediatric and neonate.
Multi-lead, 25 classifications. Asystole, VFib/VTac, Vtac, Vent. Brady, Extreme Tachy, Extreme Brady, Vrrhythm, PVCs/min, Pauses/min, Couplet, Bigeminy, Trigeminy, R on T, Run PVCs, PVC, Tachy, Brady, Missed Beats, PNP, PNC, Multif. PVC, Nonsus. Vtac, Pause, Irr. Rhythm., Afib (for adult only).

ST Segment Analysis

Intended use for adult, pediatric and neonate.

| | |
|---------------|---|
| ST range | - 2.5 to + 2.5 mV |
| ST accuracy | ± 0.02 mV or ± 10%, whichever is greater (- 0.8 to + 0.8 mV) |
| ST resolution | 0.01 mV |

QT Analysis

Intended use for adult, pediatric, and neonate.

| | |
|------------|---------------|
| Parameters | QT, QTc, ΔQTc |
|------------|---------------|

| | |
|----------------|--|
| QTc formula | Bazett, Fridericia, Framingham, or Hodges |
| QT/QTc range | 200 to 800 ms |
| QT accuracy | ± 30 ms |
| QT resolution | 4 ms |
| QTc resolution | 1 ms |
| QT-HR range | Adult: 15 to 150 bpm Pediatric/Neonate: 15 to 180 bpm |

Respiration

| | |
|---------------|--|
| Lead | I or II, auto |
| RR range | 0 to 200 rpm |
| RR accuracy | ± 1 rpm (0 to 120 rpm) ± 2 rpm (121 to 200 rpm) |
| RR resolution | 1 rpm |
| Sweep speed | 3 mm/s, 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s |
| Apnea time | 10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s |

SpO₂

Meet standards of ISO 80601-2-61.

| | |
|---------------------------|--|
| SpO ₂ module | Mindray SpO ₂ , Nellcor SpO ₂ |
| SpO ₂ range | 0 to 100 % |
| SpO ₂ accuracy | Adult/Pediatric: ± 2 % (70 to 100%) Neonate: ± 3 % (70 to 100%) |
| Perfusion indicator (PI) | Yes, for Mindray SpO ₂ |
| Pitch tone | Yes |
| Refreshing rate | ≤ 1 s |

PR

| | |
|-----------------|---|
| PR range | 20 to 300 bpm (from SpO ₂) 20 to 350 bpm (from IBP) 30 to 300 bpm (from NIBP) |
| PR accuracy | ± 3 bpm (20 to 300 bpm, from Mindray SpO ₂) ± 3 bpm (20 to 300 bpm, from Nellcor SpO ₂) ± 1 bpm or ± 1 %, whichever is greater (from IBP) ± 3 bpm or ± 3 %, whichever is greater (from NIBP) |
| Refreshing rate | ≤ 1 s |

Temperature

Meet standard of ISO 80601-2-56.

| | |
|-----------------|--------------------------------------|
| Technique | Thermal resistance |
| Channels | 2 channels |
| Temp range | 0 to 50 °C (32 to 122 °F) |
| Temp accuracy | ± 0.1 °C or ± 0.2 °F (without probe) |
| Temp resolution | 0.1 °C |
| Refreshing rate | ≤ 1 s |

NIBP

Meet standards of ISO 80601-2-30.

| | |
|---------------------------|---|
| Technique | Oscillometry |
| Operation mode | Manual, Auto, STAT, Sequence |
| Parameters | Systolic, diastolic, mean |
| Max measurement time | Adult/Pediatric: 180 s, Neonate: 90 s |
| Systolic range | Adult: 25 to 290 mmHg Pediatric: 25 to 240 mmHg Neonate: 25 to 140 mmHg |
| Diastolic range | Adult: 10 to 250 mmHg Pediatric: 10 to 200 mmHg Neonate: 10 to 115 mmHg |
| Mean range | Adult: 15 to 260 mmHg Pediatric: 15 to 215 mmHg Neonate: 15 to 125 mmHg |
| NIBP accuracy | Max mean error: ± 5 mmHg Max standard deviation: 8 mmHg |
| NIBP resolution | 1 mmHg |
| Assisting venous puncture | Yes |

IBP

Meet standard of IEC 60601-2-34.

| | |
|----------|------------|
| Channels | 2 channels |
|----------|------------|

| | | | |
|--|--|--|--|
| Sensitivity | 5 μ V/V/mmHg | ± 5 % of the reading (41 to 70 mmHg) | |
| Impedance range | 300 to 3000 Ω | ± 8 % of the reading (71 to 100 mmHg) | |
| IBP range | -50 to 360 mmHg | ± 10 % of the reading (101 to 150 mmHg) | |
| IBP accuracy | ± 1 mmHg or ± 2 %, whichever is greater | awRR range | 0 to 150 rpm |
| IBP resolution | 1 mmHg | awRR accuracy | ± 1 rpm |
| PPV range | 0 to 50 % | Data Review | |
| PAWP | Yes. | For 2G storage | |
| ICP measurement | Support | Trends data | Up to 120 hours @ 1min |
| Support waveforms overlapping. | | Events | Up to 1000 events, including parameter alarms, arrhythmia events technical alarms, and so on. |
| C.O. | | NIBP | Up to 1000 sets |
| Technique | Thermodilution | For 16G storage | |
| C.O. range | 0.1 to 20 L/min | Trends data | Up to 240 hours @ 1min, 2400 hours @ 10 min |
| C.O. accuracy | ± 0.1 L/min or ± 5 %, whichever is greater | Events | Up to 2000 events, including parameter alarms, arrhythmia events technical alarms, and so on. |
| C.O. resolution | 0.1 L/min | NIBP | Up to 3000 sets |
| TB range | 23 to 43 $^{\circ}$ C | For 2G & 16G storage | |
| TI range | 0 to 27 $^{\circ}$ C | Interpretation of resting | 20 sets of 12-lead ECG results |
| TB, TI accuracy | ± 0.1 $^{\circ}$ C (without sensor) | Full disclosure | Up to 48 hours for all parameter waveforms. The specific storage time depends on the waveforms stored and the number of stored waveforms. |
| TB, TI resolution | 0.1 $^{\circ}$ C | OxyCRG | 400 OxyCRG events |
| Artema Sidestream CO₂ | | ST review | Up to 120 hours @ 5 min |
| Meet standard of ISO 80601-2-55. | | Minitrend | Yes |
| CO ₂ sample flow rate | | Alarms | |
| 120 ml/min (DRYLINE II™ watertrap for adult/pediatric) | | Audible indicator | Yes, 3 different alarm tones, and prompt tone |
| 90/70 ml/min (DRYLINE II™ watertrap for neonate) | | Visible indicator | Red/yellow/cyan LED, and alarm message display |
| CO ₂ sample flow rate accuracy | | Provide AlarmSight infographic alarm indicator. | |
| ± 15 ml/min or ± 15 %, whichever is greater. | | Special Functions | |
| CO ₂ response time | | Clinical Assistive Application (CAA): ST Graphic™, EWS, GCS, 24h ECG summary, NIBP analysis. | |
| ≤ 5.0 s @ 120ml/min (for adult/pediatric) | | Calculations (drug, hemodynamic, Oxygenation, Ventilation, Renal), and Titration table. | |
| ≤ 4.5 s @ 90 ml/min (for neonate) | | Wi-Fi Communications | |
| ≤ 5.0 s @ 70 ml/min (for neonate) | | Protocol | IEEE 802.11a/b/g/n |
| Sweep speed | | Modulation mode | DSSS and OFDM |
| 3 mm/s, 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s | | Operating frequency | IEEE 802.11b/g/n (2.4G): ETSI/FCC/KC: 2.4 to 2.483 GHz MIC: 2.4 to 2.495 GHz IEEE 802.11a/n (5G): ETSI: 5.15 to 5.35 GHz, 5.47 to 5.725 GHz FCC: 5.15 to 5.35 GHz, 5.725 to 5.82 GHz MIC: 5.15 to 5.35 GHz KC: 5.15 to 5.35 GHz, 5.47 to 5.725 GHz, 5.725 to 5.82 GHz |
| CO ₂ range | | Channel spacing | 5 MHz @ 2.4 GHz, 20 MHz @ 5 GHz |
| 0-150 mmHg | | Wireless baud rate | IEEE 802.11a: 6 to 54 Mbps IEEE 802.11b: 1 to 11 Mbps IEEE 802.11g: 6 to 54 Mbps IEEE 802.11n: 6.5 to 72.2 Mbps |
| CO ₂ accuracy | | Output power | < 20dBm (CE requirement: detection mode- RMS) < 30dBm (FCC requirement: detection mode- peak power) |
| Full accuracy mode: | | Operating mode | Infrastructure |
| 0 - 40 mmHg: ± 2 mmHg | | Data security | WPA-PSK, WPA2-PSK, WPA-Enterprise, WPA2-Enterprise (EAP-FAST, EAP-TLS, EAP-TTLS, PEAP-GTC, PEAP-MSCHAPv2, PEAP-TLS, LEAP) |
| 41 - 76 mmHg: ± 5 % of reading | | | Encryption: TKIP and AES |
| 77 - 150 mmHg: ± 10 % of reading | | | |
| ISO accuracy mode: | | | |
| Add ± 2 mmHg to the full accuracy mode | | | |
| CO ₂ resolution | | | |
| 1 mmHg | | | |
| awRR range | | | |
| 0 to 150 rpm | | | |
| awRR accuracy | | | |
| ± 1 rpm (0 to 60 rpm) | | | |
| ± 2 rpm (61 to 150 rpm) | | | |
| Apnea time | | | |
| 10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s | | | |
| Oridion Microstream CO₂ | | | |
| Meet standard of ISO 80601-2-55. | | | |
| Sample flow rate | | | |
| 50 ^{-7.5} / ₊₁₅ ml/min | | | |
| Initialization time | | | |
| 30 s (typical) | | | |
| Response time | | | |
| 2.9 s (typical) | | | |
| Sweep speed | | | |
| 3 mm/s, 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s | | | |
| CO ₂ range | | | |
| 0 to 150 mmHg | | | |
| CO ₂ accuracy | | | |
| ± 2 mmHg (0 to 38 mmHg) | | | |
| ± 5 % of the reading (0.08 % increased in error for every 1 mmHg if the reading is more than 38 mmHg) (39 to 150 mmHg) | | | |
| awRR range | | | |
| 0 to 150 rpm | | | |
| awRR accuracy | | | |
| ± 1 rpm (0 to 70 rpm) | | | |
| ± 2 rpm (71 to 120 rpm) | | | |
| ± 3 rpm (121 to 150 rpm) | | | |
| Apnea time | | | |
| 10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s | | | |
| Capnostat Mainstream CO₂ | | | |
| Meet standard of ISO 80601-2-55. | | | |
| Rise time | | | |
| < 60 ms | | | |
| Sweep speed | | | |
| 3 mm/s, 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s | | | |
| CO ₂ range | | | |
| 0 to 150 mmHg | | | |
| CO ₂ accuracy | | | |
| ± 2 mmHg (0 to 40 mmHg) | | | |

Interfacing

| | |
|------------------|--|
| Main unit | AC power connector (1) VGA port (1) Network connector (1), RJ45 USB 2.0 connector (2) Analog output/nurse call/defib. Sync. Port (1) Equipotential grounding terminal (1) DC-in connector and docking (1) for ePM 10 |
| Barcode scanner | Support 1D and 2D barcode |
| Remote control | Support |
| Thermal recorder | 3 traces (paper 50 mm width, 20 m length) |
| Network printer | Support |

Power

| | |
|-----------------|---|
| Line voltage | 100 to 240 VAC ($\pm 10\%$) |
| Maximum current | 2.0A |
| Frequency | 50/60 Hz (± 3 Hz) |
| Battery | Rechargeable lithium-ion battery, 2600mAh/4500mAh Rechargeable smart lithium-ion battery 5600mAh ePM 10/12/15: ≥ 2 hours run time (2600mAh) ePM 10/12/15: ≥ 4 hours run time (4500mAh) ePM 10: ≥ 6 hours run time (5600mAh x1) ePM 12/15: ≥ 4.5 hours run time (5600mAh x1) ePM 12/15: ≥ 9 hours run time (5600mAh x2) |

| | |
|---------------------------|--|
| Recharge time (power off) | 2.5 hours to 90%(2600mAh) 5 hours to 90% (4500mAh) 5 hours to 90% (5600mAh x1) 10 hours to 90% (5600mAh x2) |
|---------------------------|--|

Environmental requirements

| | |
|-------------|--|
| Temperature | Operating: 0 to 40 °C Storage: -30 to 70 °C (ePM 10) Storage: -20 to 60 °C (ePM 12/15) |
| Humidity | Operating: 15 to 95 % (non condensing) Storage: 10 to 95 % (non condensing) |
| Barometric | Operating: 427.5 to 805.5 mmHg (57 to 107.4 kPa) Storage: 120 to 805.5 mmHg (16 to 107.4 kPa) |

Some of functions marked with an asterisk may not be available. Please contact your local Mindray sales representative for the most current information.

Mindray Building, Keji 12th Road South,
High-tech Industrial Park, Nanshan, Shenzhen 518057, P.R. China
Tel: +86 755 8188 8998 Fax: +86 755 26582680
E-mail: intl-market@mindray.com www.mindray.com

mindray | Healthcare with reach are registered trademarks or trademarks owned by Shenzhen Mindray Bio-medical Electronics Co., LTD
© 2018 Shenzhen Mindray Bio-medical Electronics Co., Ltd. All rights reserved. Specifications subject to changes without prior notice.
P/N: ENG-ePM 10/12/15 Datasheet-210285x4P-20190124

mindray